



PROTECTION AND COMFORT FOR YOUR HANDS

Tel: (562) 777-8026 • (888) 3ADENNA • Fax: (562) 777-8905
info@adenna.com • www.adenna.com

MATERIAL SAFETY DATA SHEET

VPF Synthetic Vinyl Powder Free Exam Gloves

Physical Data- Composition

Polyvinyl Chloride	48.54%
Di-Isononyl Phthalate (DINP)	41.26%
Di-Octyl Adipate (DOA)	4.85%
Pentanediol Diisobutyrate (TXIB)	2.43%
Epoxidized Soy Bean Oil	1.46%
Calcium-Zinc Stabilizer	1.46%

Fire & Explosion

Stability: Stable

Flash Point: N/A

Extinguishing Media: Foam, Carbon Dioxide, Water

Precaution

This product contains polyvinyl chloride (PVC). If user or a patient are allergic to vinyl or PVC or experience any discomfort, discontinue use immediately and consult with a physician.

Do not reuse gloves. Good quality exam gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.

Storage

Store in original packing in a cool, dry and well ventilated area, away from dust, sunlight, moisture, X-ray, and excessive heat above 100°F (37°C).

DISCLAIMER OF EXPRESSED AND IMPLIED WARRANTIES

Although reasonable care has been taken in the preparation of this document, we extend no warranties and make no representations as to the accuracy or completeness of the information contained therein, and assume no responsibility regarding the suitability of this information for the user's intended purposes or for the consequences of its use. Individuals should make a determination as to the suitability of the information for their particular purpose(s).